

AORTIC ENDOGRAFTS

TAA GRAFTS

Company Name	Product Name	Stent Material	Graft Material	Stent Expansion	Length (cm)	Diameter (mm)	Sheath Required for Delivery	Delivery Sheath (F)
Cook Medical	TX2	Stainless steel	Woven polyester	Self-expanding	Proximal: 108–216; distal: 127–207	28–42	No	20, 22
Medtronic Vascular	Talent Thoracic Xcelerant	Nitinol	Dacron polyester	Self-expanding	11	22, 24, 26, 28, 30, 32, 34, 36, 38, 40, 42, 44, 46	No	22, 24, 25 (OD)
W. L. Gore & Associates	Gore TAG Thoracic Endo-prosthesis	Nitinol	ePTFE	Self-expanding	10, 15, 20	26, 28, 31, 34, 37, 40	Yes	20, 22, 24

TAA ACCESSORIES: GUIDEWIRES

Company Name	Product Name	Diameter (inch)	Length (cm)	Flexible Tip Length (cm)		
Boston Scientific Corporation	Amplatz Super Stiff ST	.035, .038	145	4		
			260	1		
	Amplatz Super Stiff J	.035 .038	145, 180, 260	7		
			145, 180			
	Amplatz Super Stiff	.035 .038	145, 180, 260	6		
Meier Wire C Curve	.035	260, 300	10			
Meier Wire J Curve		185	15			
Company Name	Product Name	Diameter (inch)	Length (cm)	Primary Curve Radius (mm)	Secondary Curve Radius (mm)	Flexible Tip Length (cm)
Cook Medical	Lunderquist DC	.035	260, 300	75	15	4
	Lunderquist			0	0	

TAA ACCESSORIES: OCCLUSION/MOLDING BALLOONS

Company Name	Product Name	Size (F)	Recommended Introducer Size (F)	Guidewire Diameter (inch)	Catheter Length (cm)	Maximum Inflated Balloon Diameter (mm)
Boston Scientific Corporation	Equalizer Balloon Catheter	7	14	.038	65, 100	20, 27, 33, 40
Cook Medical	Coda Balloon Catheter	10	14	.035	120	40
Medtronic Vascular	Reliant Stent Graft Balloon Catheter	8	12	.038 or smaller	100	46