

# TAA GRAFTS AND ACCESSORIES

## TAA GRAFTS

Company	Device	Stent Material	Graft Material	Stent Expansion	Length (cm)	Diameter (mm)	Sheath Required for Delivery	Delivery Sheath (F)
Gore & Associates	TAG	Nitinol	ePTFE	Self-expanding	10, 15, 20	26, 28, 31, 34, 37, 40	Yes	20, 22, 24

## WIRE GUIDES

Company Name	Product Name	Diameter (inch)	Length (cm)	Primary Curve Radius (mm)	Secondary Curve Radius (mm)	Flexible Tip Length
Cook Incorporated	Lunderquist Extra Stiff Wire Guide (LES 3)	.035	260	75	15	4
	Lunderquist Extra Stiff Wire Guide	.035	260	0	0	4

## OCCLUSION/MOLDING BALLOONS

Company Name	Product Name	Size (F)	Recommended Introducer Size (F)	Wire Guide Diameter (inch)	Catheter Length (cm)	Maximum Inflated Balloon Diameter (mm)
Cook Incorporated	Coda Balloon Catheter	10	14	.035	120	40
Medtronic	Reliant Stent Graft Balloon Catheter	8	12	.038 or smaller	100	46

# TAA GRAFTS AND ACCESSORIES

## DIAGNOSTIC CATHETERS

Company Name	Product Name	Catheter Type: (Angiographic, sizing, or flush)	Size (F)	Recommended Introducer Size (F)	Wire Guide Diameter (inch)	Catheter Length (cm)
Cook Incorporated	Beacon Tip	Angiographic Kumpe	5	5	.035	100
	Aurous	Sizing pigtail	5	5	.035	100

## DILATORS

Company Name	Product Name	Size (F)	Dilator Length (cm)	Wire Guide Compatibility (inch)
Cook Incorporated	Hydrophilic Dilator Set	20, 24	30	.035

## EMBOLIZATION COILS

Company Name	Product Name	Extended Embolus		Coiled Embolus	
		Diameter (inch)	Length (cm)	Diameter (mm)	Configuration
Cook Incorporated	Nester	.035	14	4	11.7 loops
				6	7.4 loops
				8	5.6 loops
				10	4.5 loops
				12	3.7 loops
	Tornado	.035	4	5	
			6	6	
			8	7	
			8	8	
			8	8	
			12	10	