

OTHER DEVICES

EMBOLIZATION COILS

Company Name	Product Name	Material Used	Coil Diameter (mm)	Coil Length (cm)	Recommended Delivery Catheter Size Inner Diameter (inch)	Indicated Use		
Boston Scientific Corporation*	VortX-18 Diamond Shape Fibered Platinum Coils	Platinum	Small secondary/large secondary: 2/3, 2/4, 2/5, 2/6	2.3, 4.1, 5.8, 8	.021	.018-inch fibered platinum coils are intended for arterial and venous embolizations in the peripheral vasculature		
	VortX-18 Fibered Platinum Coils		Small secondary/large secondary: 2/3, 2/4, 2/5, 2/6	2.2, 4.2, 6, 8.5				
	Straight .018-inch Fibered Platinum Coils		Straight	.2, .5				
	Figure-8 .018-inch Fibered Platinum Coils		2	1				
	Multi-Loop .018-inch Fibered Platinum Coils		4	3				
	Complex Helical .018-inch Fibered Platinum Coils		4, 6, 7, 8, 10, 11	2, 3, 4, 5, 6, 7				
	Multi-Loop .035-inch Fibered Platinum Embolization Coils		3, 4, 5, 6, 7, 9	2, 3, 4, 5, 6			.038	For arterial and venous embolizations in the peripheral vasculature
	VortX-35 Vascular Occlusion Coils		Apex diameter/base diameter: 2/4, 3/5, 3/6, 3/7	3, 3.5, 5.3, 6.7				
Interlock Fibered IDC Occlusion System	Platinum-tungsten alloy	2D Helical: 3, 4, 5, 6, 8, 10, 12, 14	6, 8, 10, 12, 15, 20, 30	.021	To obstruct or reduce rate of blood flow in the peripheral vasculature; device is not intended for neurovascular use			
		VortX Diamond: small secondary/large secondary 2/3, 2/4, 2/5, 2/6	2.3, 4.1, 5.8, 8					

*Please reference appropriate product instructions for use from Boston Scientific for a list of indications, warnings, precautions, and potential adverse effects.

OTHER DEVICES

EMBOLIZATION COILS (CONTINUED)

Company Name	Product Name	Material Used	Coil Diameter (mm)	Coil Length (cm)	Recommended Delivery Catheter Size Inner Diameter (inch)	Indicated Use
Cook Medical	Nester Embolization Coils	Platinum	4, 6, 8, 10, 12	14	.035	Used for arterial and venous embolization
	MicroNester Embolization Coils		3, 4, 6, 8, 10		.018	
	Tornado Embolization Coils		5-3, 6-3, 7-3, 8-4, 8-5, 10-5	4.1–12.5	.035	Used for embolization of selective vessel supply to arteriovenous malformations, trauma, bleeding, and other vascular lesions
	Tornado Embolization Microcoils		3-2, 4-2, 5-2, 6-2, 7-3, 8-4, 10-4	2–14.2	.018	
	Hilal Embolization Coils		Straight to 5	.5–3		Used for embolization of selective vessel supply to arteriovenous malformations and other vascular lesions of the brain, spinal cord, and spine
	MReye Embolization Coils	Inconel	3, 4, 5, 8, 10, 12, 15	2, 3, 4, 5, 8, 15	.035, .038	Used for arterial and venous embolization
	Stainless Steel Embolization Coils	Stainless steel	6, 8, 10, 15, 20	8, 10, 15	.052	
	Flipper Detachable Embolization Coils		3, 5, 6.5, 8	3, 4, 5, 6, 8, 10, 12	.041	
Terumo Interventional Systems	AZUR Peripheral HydroCoil Pushable .018-inch System	Platinum and hydrogel	2, 3, 4, 5, 6, 8, 10	2, 4, 6, 10, 14	.022	
	AZUR Peripheral HydroCoil Pushable .035-inch System		4, 5, 6, 8, 10, 15	4, 6, 10, 14	.035	
	AZUR Peripheral HydroCoil Detachable .018-inch System		4, 6, 8, 10, 15, 20	10, 15, 20	.022–.025	